



O₂ High-Flow Oxygen Therapy
When comfort is efficient

Oxygen therapy administered through nasal cannulas or masks has always been the first approach to alleviate hypoxemia in patients.

A new feature available in our ventilators

In conventional oxygen therapy, the flow used is generally low, due to the fact that high flow rates without an appropriate humidity level may cause discomfort. Within this context, FiO₂ received by the patient may also vary, depending greatly on the patient's inspiratory flow rate. In this scenario, high-flow nasal cannula (HFNC) allow oxygen administration of up to 80 L/min with a high humidity level; solving the issues raised in the past. Advance, Neo and Ts* ventilators now have this feature, which allows an optimal transition between Invasive Ventilation, Non-Invasive Ventilation and High-Flow Oxygen Therapy.

Physiological Benefits

High Humidity: The Key to Tolerance

Dry and cold gases administered through the nasal cavity produces irritation and drying of the nasal airway mucosa, which reduces the tolerance to the method and increases the mucociliary clearance deterioration.

Using high humidity level HFNC allows toleration of the administered flows and prevents the impairment of the cleaning and defense mechanisms of the airways. During the weaning period, it alleviates the larynx and trachea inflammation, preventing the postextubation failure.

CO₂ Clearance

During HFNC application, a reduction in the respiratory rate has been observed, as well as a reduction of the minute volume without PaCO₂ increase, which suggests a more efficient ventilation, possibly originated by the CO₂ clearance of the upper airway and reduction of dead space.

Controlled and predictable FiO₂

In conventional oxygen therapy, FiO₂ is highly variable, depending greatly on the patient's inspiratory flow. Through the administration of flows that exceed the inspiratory peak flow, HFNC guarantees continuous and controlled FiO₂.

CPAP-like Effect

Although HFNC is an open system, the stream flow it uses generates expiratory resistance, which creates a certain



level of pressure in the upper airways, equivalent to CPAP levels of up to 5 cmH₂O with a similar physiological effect.

Uses and Application

HFNC has been used for some years in practice, and the number of application areas has been increasing:

- Handling pre and postextubation periods.
- Hypoxemic and hypercapnic respiratory failure.
- Heart failure.
- Sleep apnea.
- As an alternative to Non-Invasive Ventilation.
- In handling pediatric and neonatal patients.

*As of software version N11-01.14

50 years of innovation and development in mechanical ventilators.

www.tecmeglobal.com

Córdoba

Av. Circunvalación (N-0) - Agustín Tosco 3040
X5008HJY - Córdoba, Argentina
Tel.: +54 351 4144600 - Fax: +54 351 4144605
e-mail: info@tecme.com.ar

Buenos Aires

Trole 557 - 2º Piso, Nueva Pompeya, C.A.B.A.,
C1437DKK - Buenos Aires, Argentina
Tel.: +54 11 1561387326
e-mail: info@tecme.com.ar

Norcross

2825 Pacific Drive, Suite B
Norcross, GA 30071 - USA
P: +1 770 409 9172 - F: +1 770 729 8176
e-mail: info@tecme-usa.com